

Sirolimus-Eluting Stents for the Treatment of Stenoses in Small Coronary Arteries: What Have We Learned?

Ricardo A. Costa, MD and Alexandra J. Lansky, MD

Cardiovascular Research Foundation and Columbia University Medical Center - Nova York, NY, USA

Percutaneous coronary intervention (PCI) in small vessels (broadly defined as stenosis in vessel with reference diameter <2.75-2.80mm, or <3.00mm)^{1,2}, have been historically associated with high rates of restenosis (18-52%) and vessel revascularization (up to 27%)^{1,3-5}. Prior to drug-eluting stents (DES), a strong correlation between restenosis and vessel size was recognized, with an inverse association between vessel size and angiographic restenosis; this was attributed to the disproportionately greater amount of neointimal hyperplasia (NIH) relative to the vessel caliber in smaller vessels^{1,3}. This phenomenon seemed to be aggravated in diabetics who often present with "small" arteries because of diffuse coronary disease and exaggerated neointimal proliferative response⁴. Several studies had attempted to demonstrate the efficacy of stenting versus balloon angioplasty in small vessels, but results have been variable. A recent meta-analysis of 11 randomized trials (2,971 patients) comparing bare metal stents (BMS) to balloon angioplasty in vessels 2.22-2.60mm, demonstrated lower restenosis with by stent implantation ($p=0.003$), with a associated relative risk of 0.77 (95% confidence interval 0.65, 0.92)¹.

Since the introduction of sirolimus-eluting stents (SES), many expected that the similar effectiveness of SES for preventing restenosis in uncomplicated lesions would extend to high risk subsets including patients with small vessels. In this issue of the *Arquivos Brasileiros de Cardiologia*, Devito et al⁶ reports the outcomes of 80 patients enrolled in a prospective, non-randomized comparison of SES (Cypher[®] stent) versus a BMS (ML-Pixel[®] stent), for the treatment of *de novo* coronary lesions in vessels 2.20-2.75mm in diameter. All stents were implanted successfully and final luminal dimensions were similar in both groups. (It is important to note that in this investigation, patients were consecutively enrolled at two different time points, and even though baseline clinical demographics were comparable, the two groups were not perfectly matched according to some important

criteria, including lesion complexity and stent size.) At follow-up, late lumen loss (LLL), restenosis and NIH area were significantly reduced with SES, and target lesions revascularization (TLR) was 3.5-fold lower in SES versus BMS patients, a difference that did not reach statistical significance.

The performance of SES in small vessels has been evaluated in several randomized clinical trials and registries (Table 1). LLL, a surrogate of NIH, was significantly decreased with SES compared to BMS in "small vessels" by 98.7%, 78.8%, 81%, 88.2%, and 82.2% in RAVEL, SIRIUS, E-SIRIUS, C-SIRIUS, and SES-SMART, respectively^{2,5,7-10}. In Devito's study, LLL was decreased by 77.5%, $p<0.001$ ⁶. A slightly higher in-stent LLL (0.36mm) was found in SIRIUS-2.25 compared to the other studies (0.01-0.21mm) (Table 1). SIRIUS-2.25 had increased rates of diabetics (40%) and more complex lesions (71% B2/C)¹¹. Diabetics have a higher LLL compared to non-diabetics^{12,13}. In Devito's analysis, which also had 40% of diabetes but less complex lesions, LLL was 0.25mm⁶. Overall, these LLL were comparable to the LLL found in lesions in larger vessels treated with SES^{7-9,14}, and were considerably lower compared to BMS (0.80-1.05mm)^{2,5,7,10}, confirming the consistency and effectiveness of SES in preventing NIH, despite differences in rates of diabetes, vessel size and lesion characteristics.

The classic inverse relationship between vessel size and restenosis was not seen with SES in RAVEL (0% restenosis in all vessel sizes)⁷. But, in a sub-analysis of the SIRIUS trial (which had longer and more complex lesions), lesions were divided into tertiles according to vessel size, and in-stent restenosis was 5%, 2.5%, and 1.9% with SES in small (2.32mm), "medium" (2.78mm), and large (3.31mm) vessels. Conversely, in-segment restenosis was increased in small vessels (17.6%), compared to "medium" (6.6%), and large (1.9%) vessels⁹. As a result, the risk reductions for in-segment restenosis with SES decreased by 93.7%, to 81.7%, to 58.8% in large,

Table 1 - SES In "Small Vessel" Trials

	RAVEL ¹⁷	SIRIUS ¹¹⁷	E-SIRIUS ¹⁰	C-SIRIUS ⁹	SES-SMART ²	RESEARCH ¹⁴	SVELTE ¹⁵	SIRIUS 2.25mm ¹¹	Devito et al ⁶
Number of patients	42	350	175	50	129	91 [§]	101	100	50
Diabetics, %	19	27.1	19	24	19.4	26	26.7	40	40
Lesion class B2/C (ACC/AHA), %	61.9	59 [§]	-	64	30.4	92 [†]	-	71	32
QCA									
RD, mm	2.09	2.41	2.60	2.65	2.22	1.88	2.36	2.04	2.44
Lesion length, mm	9.5	14.6	14.9	14.5	13.0	12.3	14.5	12.2	13.8
Follow-up (6-8 months)									
Late lumen loss, mm									
In-stent [#]	0.01	0.17	0.20	0.12	0.16	-	0.21	0.36	0.25
In-segment ^{**}	-0.04	0.23	0.19	0.12	0.16	0.07	0.21	0.23	0.30
Binary restenosis, %									
In-stent [#]	0	3.9	3.9	0	4.9	10.7	3.2	11.7	0
In-segment ^{**}	0	11.6	5.9	2.3	9.8	10.7	6.3	16.9	4
IVUS									
NIH area, mm ²	0.11	0.58	-	-	-	-	0.08	0.16	0.24
Clinical outcome									
Follow-up	12-month	12-month	9-month	9-month	8-month	12-month	12-month	6-month	8-month
TLR, %	0	5.7	4	4	7	5.5	3	4.3	4
Any MACE, ^{††} %	5.8 ^{‡‡}	9.4	8	4	9.3 ^{§§}	7.7	7.9	7.4	4
SAT (n)	0	1	2	1	1	1	0	1	0
<p>Values are expressed as mean or frequencies (% of column total). *Sub-analysis reporting the lowest tertile of vessel size. †SES sub-group from SIRIUS "matched" (and compared) with SVELTE by diabetes, RD (2.20-2.75mm) and lesion length (>15mm, estimated). ‡Sub-group with lesions treated with 2.25mm SES only. §Includes 112 lesions. Type B2 only. ¶Overall percentage of patients with complex lesions in the study. #In-stent – defined as within the stented segment. **In-segment – defined as spanning the stented segment plus the 5 mm proximal and distal peri-stent areas. ††Major adverse clinical events, typically defined as death, myocardial infarction and/or target lesion revascularization. ‡‡Overall percentage of patients with MACE in the SES group in RAVEL. §§Includes 1 (0.8%) CVA. ACC/AHA=American College of Cardiology/American Heart Association; CVA=cerebral-vascular accident; IVUS=intravascular ultrasound; QCA=quantitative coronary angiography; RD=reference (vessel) diameter; SAT=subacute stent thrombosis; TLR=target lesion revascularization</p>									

"medium", and small vessels, respectively. In SIRIUS-2.25, restenosis rates were comparable to those of the SIRIUS "small vessel" tertile, and were associated with stent length, diabetes and number of stents implanted¹¹. In Devito's report, there was no in-stent restenosis and only 4% in-segment restenosis with SES compared to 33.3% in-stent and 36.7% in-segment restenosis with BMS ($p < 0.001$)⁶. Similar results were found in RAVEL, E- and C-SIRIUS, and SVELTE^{5,7,10,15}. In RESEARCH, vessel size was considerably smaller compared to the other studies (Table 1), and restenosis (only in-stent) occurred mostly after treatment of ostial lesions¹⁴. Importantly, in the SES-SMART sub-analysis in diabetics, SES was associated with significant decrease in LLL and restenosis compared to BMS; however, when considering only insulin dependent diabetics (35%), restenosis with SES rose to 40%¹². These data convey important considerations regarding SES in small vessels, including: 1) SES of small vessels have higher restenosis rates compared to SES of larger vessels; 2) restenosis appears to increase according

to lesion severity and stented length; 3) diabetics may experience higher rates of restenosis. This may be related not only to the hyperproliferative and aggressive vascular response often observed, but also to suboptimal stent deployment (incomplete lesion coverage, inadequate stent expansion, see below), since diabetics frequently present with long and diffuse disease (as mentioned before); and 4) the higher "in-segment" restenosis (compared to "in-stent") observed in the majority of studies (Table 1) may be reflecting incomplete lesion coverage and/or injured segment during PCI. This situation has been recognized^{8,16} and improvement in the "DES" technique – minimize peri-lesion trauma using short pre-dilatation balloons, complete lesion/injured segment coverage, short post-dilatation balloons (positioned within the stented area to prevent injury outside of the stent edges), and overlap multiple stents avoiding gaps, may be critical to optimize results in small vessels.

Previous studies with IVUS demonstrated that final luminal dimensions in vessels <2.75mm predict vessel

revascularization¹⁷. Similar results were found with SES. In the IVUS sub-study of the SIRIUS trial, a final minimum stent area >4.5mm² for vessels <2.8mm (by QCA) was found to be a threshold that predicted an “adequate” IVUS lumen at follow-up (>4.0mm²). The positive predictive value of the IVUS stent dimensions was 90%¹⁸. In addition, Takebayashi et al reported a series of patients with target vessel failure post-SES implantation where the majority of failures (especially in-stent restenosis) were associated with stent *underexpansion*¹⁹. This is not entirely unexpected as once an effective drug, (in this case, sirolimus), inhibited most of the NIH (Table 1), the main cause of in-stent restenosis became stent *underexpansion*. At last, a larger reference percentage of plaque area and a larger edge stent area/reference minimum lumen area were associated with edge stenosis with SES in SIRIUS²⁰. These evidences suggest that successful SES is relied upon a combination of SES’ ability to prevent NIH formation and optimal stenting technique. In the context of small vessels, IVUS guided PCI for SES implantation should be strongly considered and even recommended, as it: 1) allows accurate assessment of vessel size and lesion length before the procedure, helping operators to select the proper device

and technical approach to achieve full lesion coverage, and proper stented segment matching; and 2) may identify inadequate stent expansion, permitting the operator to post dilate the stent to achieve optimal stent expansion.

Finally, implantation of SES in small vessels has been associated with a high procedural success rate (>95%) and sustained safety (comparable to larger vessels treated with SES), with low rates of stent thrombosis and MACE during the mid- and long term follow-up (Table 1)^{2,5,6,8,10,11,14,15}. In addition, SES led to a 70% to 81% relative reduction of TLR compared to BMS in “small vessel” randomized trials^{2,5,10}. Importantly, both TLR and MACE rates with SES (Table 1) were remarkably low in all studies (<10%), especially given the high-risk lesions treated and the high-risk population. These results confirm the dramatic impact and overall benefits of SES compared to BMS and historical controls of PCI in small vessels. Nevertheless, despite the impressive progress obtained in the field, small vessel is still predictive of restenosis after SES implantation⁸; therefore, continued efforts, including optimize DES technique and development of novel DES technologies specifically designed for small vessels, are still needed to improve outcome.

REFERENCES

- Moreno R, Fernandez C, Alfonso F, Hernandez R, Perez-Vizcayno MJ, Escaned J et al. Coronary stenting versus balloon angioplasty in small vessels: a meta-analysis from 11 randomized studies. *J Am Coll Cardiol*. 2004;43:1964-72.
- Ardissino D, Cavallini C, Bramucci E, Indolfi C, Marzocchi A, Manari A et al. Sirolimus-eluting vs uncoated stents for prevention of restenosis in small coronary arteries: a randomized trial. *JAMA*. 2004;292:2727-34.
- Holmes DR Jr, Kereiakes DJ. The approach to small vessels in the era of drug-eluting stents. *Rev Cardiovasc Med*. 2005;6 Suppl 1:S31-7.
- Saucedo JF, Popma JJ, Kennard ED, Talley JD, Lansky A, Leon MB et al. Relation of coronary artery size to one-year clinical events after new device angioplasty of native coronary arteries (a New Approach to Coronary Intervention [NACI] Registry Report). *Am J Cardiol*. 2000;85:166-71.
- Schampaert E, Cohen EA, Schluter M, Reeves F, Traboulsi M, Title LM et al. The Canadian study of the sirolimus-eluting stent in the treatment of patients with long de novo lesions in small native coronary arteries (C-SIRIUS). *J Am Coll Cardiol*. 2004;43:1110-5.
- Devito FS, Sousa AGMR, Feres F, Abizaid A, Staico R, Mattos LAP et al. Comparative Analysis of Intimal Hyperplasia After Sirolimus-Eluting Stent and Thin-Strut Bare Metal Stent Implantation in Small Coronary Arteries. *Arq Bras Cardiol*. 2006;86:268-75.
- Regar E, Serruys PW, Bode C, Holubarsch C, Guernonprez JL, Wijns W et al. Angiographic findings of the multicenter Randomized Study With the Sirolimus-Eluting Bx Velocity Balloon-Expandable Stent (RAVEL): sirolimus-eluting stents inhibit restenosis irrespective of the vessel size. *Circulation*. 2002;106:1949-56.
- Moses JW, Leon MB, Popma JJ, Fitzgerald PJ, Holmes DR, O’Shaughnessy C et al. Sirolimus-eluting stents versus standard stents in patients with stenosis in a native coronary artery. *N Engl J Med*. 2003;349:1315-23.
- Popma JJ, Leon MB, Moses JW, Holmes DR Jr, Cox N, Fitzpatrick M et al. Quantitative assessment of angiographic restenosis after sirolimus-eluting stent implantation in native coronary arteries. *Circulation*. 2004;110:3773-80.
- Schofer J, Schluter M, Gershlick AH, Wijns W, Garcia E, Schampaert E et al. Sirolimus-eluting stents for treatment of patients with long atherosclerotic lesions in small coronary arteries: double-blind, randomised controlled trial (E-SIRIUS). *Lancet*. 2003;362:1093-9.
- Nikolsky E, Moses JW, Cambier P, Bachinsky B, O’Shaughnessy C, Mehran R et al. Results from the 2.25mm Sirolimus-Eluting BS Velocity Stent Registry in Patients with Native Coronary Artery Lesions. *Am J Cardiol*. 2005;96 (Suppl 7A):179H.
- Ortolani P, Ardissino D, Cavallini C, Bramucci E, Indolfi C, Aquilina M et al. Effect of sirolimus-eluting stent in diabetic patients with small coronary arteries (a SES-SMART substudy). *Am J Cardiol*. 2005;96:1393-8.
- Moussa I, Leon MB, Baim DS, O’Neill WW, Popma JJ, Buchbinder M et al. Impact of sirolimus-eluting stents on outcome in diabetic patients: a SIRIUS (SIROLImUS-coated Bx Velocity balloon-expandable stent in the treatment of patients with de novo coronary artery lesions) substudy. *Circulation*. 2004;109:2273-8.
- Lemos PA, Arampatzis CA, Saia F, Hoye A, Degertekin M, Tanabe K et al. Treatment of very small vessels with 2.25-mm diameter sirolimus-eluting stents (from the RESEARCH registry). *Am J Cardiol*. 2004;93:633-6.
- Sousa JE. SVELTE: Multicenter, Controlled Study of Sirolimus-Eluting Stents in Small Vessels. *Transcatheter Cardiovascular Therapeutics (TCT)*, Washington, DC [presentation]. http://www.tctmd.com/cportal/appmanager/tctmd/main?_nfpb=true&_pageLabel=TCTMDContent&hdCon=814312. 2004.
- Lemos PA, Saia F, Ligthart JM, Arampatzis CA, Sianos G, Tanabe K et al. Coronary restenosis after sirolimus-eluting stent implantation: morphological description and mechanistic analysis from a consecutive series of cases. *Circulation*. 2003;108:257-60.
- Iakovou I, Mintz GS, Dangas G, Abizaid A, Mehran R, Lansky AJ et al. Optimal final lumen area and predictors of target lesion